儀寶電子股份有限公司

I PAO ELECTRONICS CO., LTD

校 正 報 告 書

REPORT OF CALIBRATION

Report No.： **D619P081**　　校正日期（Date）：**24.Jun.2014**

申請者：Applicant	麥司儀器有限公司	儀器名稱：Equipment	真空壓力計
製造商：Manufacturer	**ATLANTIS**	型號：Model No.	**-760~0 mmHg**
序號：Serial No.	**130103007**		
申請者地址：Applicant address	新北市新莊區五權一路 7 號 6 樓 607 室		

校正時使用之工作標準器
Working Standards

儀器名稱 Equipment	製造商/型號 MFG/Model No.	識別號碼 I.D. No.	校正機構 Cal.Sources	報告號碼 Report.No.	校正日期 Cal. Date	有效日期 Due. Date
數字型壓力計	SI/PC6-PRO-0035-C-3	7076	TAF(1805)	14A043098	14.Mar.2014	13.Mar.2015

追溯源
Calibration sources

儀器名稱 Equipment	製造商/型號 MFG/Model No.	識別號碼 I.D. No.	校正機構 Cal.Sources	報告號碼 Report.No.	校正日期 Cal. Date
數字型壓力計	SI/PC6-PRO-0035-C-3	7076	TAF(1805)	14A043098	14.Mar.2014

儀寶電子股份有限公司特此證明本報告書內之受校儀器已與上列標準做過比較校正，用以校正之標準器可追溯至國家度量衡標準實驗室。本報告僅對送校儀器之校正項目有效。本報告不可摘錄部份複製無效。

IPE Ltd here by certifies that equipment noted here in has been compared with the above listed standards. The standards used to perform this calibration are traceable to NML. This calibration report is valid only to the items calibrated. Reproduced calibration report in partial is not effective.

I PAO ELECTRONICS CO LTD. 儀寶電子 儀器校驗中心 校正專用 新北市新莊區五權一路76號6F

實驗室主管 Laboratory Manager 陳謙毅

報告簽署人 Report Signatory Thomas

校正報告書
REPORT OF CALIBRATION

Report No. **D619P081**

1.校正結果

量測值(mmHg)	器示值(mmHg)	器差值(mmHg)
0.0	0.0	0.0
- 101.7	-100.0	1.7
- 301.4	-300.0	1.4
- 501.6	-500.0	1.6
- 601.4	-600.0	1.4
- 701.3	-700.0	1.3

2.校正說明：

2.1 校正環境：

2.1.1 溫度為（23±2）℃

2.1.2 相對濕度為（50±15）%

2.2 校正方法為依據本公司之壓力錶校正程序手冊(非 LOGO)(PCP-003-NL)

2.3 量測值：標準實際產生之訊號之值。

2.4 器示值：待校正之儀器所顯示之值。

2.5 器差值＝器示值－量測值

2.6 壓力之 SI UNIT 為 Pa，1 Pa=1.0197×10^{-5} kgf/cm^2=0.00014504 psi 。